FUGITIVE VOC EMISSION SURVEY
SCREENING REPORT

First QUARTER 2012

Norlite Corporation
628 South Saratoga Street
Cohoes, NY 12047

APPENDIX A

Microtip Calibration Data Sheets

MICROTIP CALIBRATION DATA SHEET

Microtip Serial #: PPWM0051

#1 Gas Standard (type): Air, zero **SN:** 0129005-001 **Certified Conc.:** zero air

#2 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc.:** 1001 ppm

#3 Gas Standard (type): **SN:** **Certified Conc.:**

CALIBRATION DATE[1]:	CALIBRATOR	RESPONSE	RECALIBRATED?	COMMENTS
1/24/2012	TVV	Gas #1 0 Gas #2 1001 Gas #3	N N	Calibrated
1/25/2012	TVV	Gas #1 0 Gas #2 1000 Gas #3	N N	Calibrated
1/26/2012	TVV	Gas #1 0 Gas #2 999 Gas #3	N N	Calibrated
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		

RESPONSE TIME CHECK[2]

Gas Standard SN: BAJ-289-990-1

Date: 1/24/2012

Conc: 1001

Time Response (sec.)	Instrument Response (ppm)	Comments
4	1000	
3	999	
3	1000	
Average Response Time	3.33 seconds	

[1] Calibration is required at the beginning of each sampling day.

[2] Response time check is required once per month.

40 CFR 60, APPENDIX A, METHOD 21
CALIBRATION PRECISION TEST DATA SHEET
(QUARTERLY)

Microtip Serial #: PPWM0051

#1 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc.:** 1001 ppm

#2 Gas Standard (type): _____ **SN:** _____ **Certified Conc.:** _____

#3 Gas Standard (type): _____ **SN:** _____ **Certified Conc.:** _____

DATE	INITIALS	METER READINGS	AVERAGE DIFFERENCE[1]	CALIBRATION PRECISION %[2]	COMMENTS
1/24/2012	TVV	#1 Gas- a 998 b 1000 c 1003	2.00	0.20	
		#2 Gas- a b c			
		#3 Gas- a b c			
1/25/2012	TVV	#1 Gas- a 1002 b 997 c 1000	2.00	0.20	
		#2 Gas- a b c			
		#3 Gas- a b c			
1/26/2012	TVV	#1 Gas- a 996 b 997 c 1000	3.33	0.33	
		#2 Gas- a b c			
		#3 Gas- a b c			
		#1 Gas- a b c			
		#2 Gas- a b c			
		#3 Gas- a b c			

[1] Average difference is done by calculating difference between each meter reading and certified concentration, divided by the number of meter readings.

[2] Calibration precision percentage is obtained by dividing value in step (1) by the certified concentration and multiplying by 100.

FUGITIVE VOC EMISSION SURVEY
SCREENING REPORT

Second Quarter 2012

Norlite Corporation
628 South Saratoga Street
Cohoes, NY 12047

APPENDIX A

Microtip Calibration Data Sheets

40 CFR 60, APPENDIX A, METHOD 21
CALIBRATION PRECISION TEST DATA SHEET
(QUARTERLY)

Microtip Serial #: PPWM0051

#1 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc.:** 1001 ppm

#2 Gas Standard (type): _____ **SN:** _____ **Certified Conc.:** _____

#3 Gas Standard (type): _____ **SN:** _____ **Certified Conc.:** _____

DATE	INITIALS	METER READINGS	AVERAGE DIFFERENCE[1]	CALIBRATION PRECISION %[2]	COMMENTS
4/23/2012	TVV	#1 Gas- a 995 b 997 c 999	4.00	0.40	
		#2 Gas- a b c			
		#3 Gas- a b c			
4/24/2012	TVV	#1 Gas- a 1003 b 998 c 998	2.67	0.27	
		#2 Gas- a b c			
		#3 Gas- a b c			
		#1 Gas- a b c			
		#2 Gas- a b c			
		#3 Gas- a b c			
		#1 Gas- a b c			
		#2 Gas- a b c			
		#3 Gas- a b c			

[1] Average difference is done by calculating difference between each meter reading and certified concentration, divided by the number of meter readings.

[2] Calibration precision percentage is obtained by dividing value in step (1) by the certified concentration and multiplying by 100.

MICROTIP CALIBRATION DATA SHEET

Microtip Serial #: PPWM0051

#1 Gas Standard (type): Air, zero **SN:** 0129005-001 **Certified Conc.:** zero air

#2 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc.:** 1001 ppm

#3 Gas Standard (type): **SN:** **Certified Conc.:**

CALIBRATION DATE[1]:	CALIBRATOR	RESPONSE	RECALIBRATED?	COMMENTS
4/23/2012	TVV	Gas #1 0 Gas #2 1001 Gas #3	N N	Calibrated
4/24/2012	TVV	Gas #1 0 Gas #2 1000 Gas #3	N N	Calibrated
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		

RESPONSE TIME CHECK[2]

Gas Standard SN: BAJ-289-990-1 Date: 4/23/2012 Conc: 1001

Time Response (sec.)	Instrument Response (ppm)	Comments
4	998	
4	999	
3	1001	
Average Response Time	3.33 seconds	

[1] Calibration is required at the beginning of each sampling day.

[2] Response time check is required once per month.

FUGITIVE VOC EMISSION SURVEY
SCREENING REPORT

Third QUARTER 2012

Norlite Corporation
628 South Saratoga Street
Cohoes, NY 12047

APPENDIX A

Microtip Calibration Data Sheets

MICROTIP CALIBRATION DATA SHEET

Microtip Serial #: PPWM0051

#1 Gas Standard (type): Air, zero **SN:** 0129005-001 **Certified Conc.:** zero air

#2 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc.:** 1001 ppm

#3 Gas Standard (type): **SN:** **Certified Conc.:**

CALIBRATION DATE[1]:	CALIBRATOR	RESPONSE	RECALIBRATED?	COMMENTS
7/16/2012	TVV	Gas #1 0 Gas #2 1002 Gas #3	N N	Calibrated
7/17/2012	TVV	Gas #1 0 Gas #2 999 Gas #3	N N	Calibrated
7/24/2012	TVV	Gas #1 0 Gas #2 1001 Gas #3	N N	Calibrated
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		

RESPONSE TIME CHECK[2]

Gas Standard SN: BAJ-289-990-1 Date: 7/16/2012 Conc: 1001

Time Response (sec.)	Instrument Response (ppm)	Comments
3	997	
4	1000	
4	1000	
Average Response Time (sec)	3.67	

[1] Calibration is required at the beginning of each sampling day.
[2] Response time check is required once per month.

40 CFR 60, APPENDIX A, METHOD 21
CALIBRATION PRECISION TEST DATA SHEET
(QUARTERLY)

Microtip Serial #: PPWM0051

#1 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc. (ppm):** 1001

#2 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

#3 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

DATE	INITIALS		METER READINGS	AVERAGE DIFFERENCE[1]	CALIBRATION PRECISION %[2]	COMMENTS
7/16/2012	TVV		#1 Gas-			
		a	994	4.33	0.43	
		b	996			
		c	1000			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
7/17/2012	TVV		#1 Gas-			
		a	1002	0.67	0.07	
		b	1001			
		c	1002			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
7/24/2012	TVV		#1 Gas-			
		a	998	2.33	0.23	
		b	999			
		c	999			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
			#1 Gas-			
		a		1001.00	100.00	
		b				
		c				
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			

[1] Average difference is done by calculating difference between each meter reading and certified concentration, divided by the number of meter readings.

[2] Calibration precision percentage is obtained by dividing value in step (1) by the certified concentration and multiplying by 100.

FUGITIVE VOC EMISSION SURVEY
SCREENING REPORT

Fourth Quarter 2012

Norlite Corporation
628 South Saratoga Street
Cohoes, NY 12047

APPENDIX A

Microtip Calibration Data Sheets

MICROTIP CALIBRATION DATA SHEET

Microtip Serial #: PPWM0051

#1 Gas Standard (type): Air, zero **SN:** 0129005-001 **Certified Conc.:** zero air

#2 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc.:** 1001 ppm

#3 Gas Standard (type): **SN:** **Certified Conc.:**

CALIBRATION DATE[1]:	CALIBRATOR	RESPONSE	RECALIBRATED?	COMMENTS
10/22/2012	TVV	Gas #1 0 Gas #2 999 Gas #3	N N	Calibrated
10/23/2012	TVV	Gas #1 0 Gas #2 1001 Gas #3	N N	Calibrated
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		

RESPONSE TIME CHECK[2]

Gas Standard SN: BAJ-289-990-1 Date: 10/22/2012 Conc: 1001

Time Response (sec.)	Instrument Response (ppm)	Comments
3	999	
3	1001	
3	1000	
Average Response Time	3.00 seconds	

[1] Calibration is required at the beginning of each sampling day.
[2] Response time check is required once per month.

40 CFR 60, APPENDIX A, METHOD 21
CALIBRATION PRECISION TEST DATA SHEET
(QUARTERLY)

Microtip Serial #: PPWM0051

#1 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc.:** 1001 ppm

#2 Gas Standard (type): _____ **SN:** _____ **Certified Conc.:** _____

#3 Gas Standard (type): _____ **SN:** _____ **Certified Conc.:** _____

DATE	INITIALS	METER READINGS	AVERAGE DIFFERENCE[1]	CALIBRATION PRECISION %[2]	COMMENTS
10/22/2012	TVV	#1 Gas- a 997 b 1000 c 1002	2.00	0.20	
		#2 Gas- a b c			
		#3 Gas- a b c			
10/23/2012	TVV	#1 Gas- a 1002 b 1002 c 1000	1.00	0.10	
		#2 Gas- a b c			
		#3 Gas- a b c			
		#1 Gas- a b c			
		#2 Gas- a b c			
		#3 Gas- a b c			
		#1 Gas- a b c			
		#2 Gas- a b c			
		#3 Gas- a b c			

[1] Average difference is done by calculating difference between each meter reading and certified concentration, divided by the number of meter readings.

[2] Calibration precision percentage is obtained by dividing value in step (1) by the certified concentration and multiplying by 100.

FUGITIVE VOC EMISSION SURVEY
SCREENING REPORT

First QUARTER 2013

Norlite, LLC
628 South Saratoga Street
Cohoes, NY 12047

APPENDIX A

Microtip Calibration Data Sheets

40 CFR 60, APPENDIX A, METHOD 21
CALIBRATION PRECISION TEST DATA SHEET
(QUARTERLY)

Microtip Serial #: PPWM0051

#1 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc. (ppm):** 1001

#2 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

#3 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

DATE	INITIALS		METER READINGS	AVERAGE DIFFERENCE[1]	CALIBRATION PRECISION %[2]	COMMENTS
1/17/2013	TVV		#1 Gas-			
		a	998	2.67	0.27	
		b	998			
		c	999			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
1/18/2013	TVV		#1 Gas-			
		a	998	1.33	0.13	
		b	1002			
		c	1001			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
			#1 Gas-			
		a		1001.00	100.00	
		b				
		c				
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
			#1 Gas-			
		a		1001.00	100.00	
		b				
		c				
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			

[1] Average difference is done by calculating difference between each meter reading and certified concentration, divided by the number of meter readings.

[2] Calibration precision percentage is obtained by dividing value in step (1) by the certified concentration and multiplying by 100.

MICROTIP CALIBRATION DATA SHEET

Microtip Serial #: PPWM0051

#1 Gas Standard (type): Air, zero **SN:** 0129005-001 **Certified Conc.:** zero air

#2 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc.:** 1001 ppm

#3 Gas Standard (type): **SN:** **Certified Conc.:**

CALIBRATION DATE[1]:	CALIBRATOR	RESPONSE	RECALIBRATED?	COMMENTS
1/17/2013	TVV	Gas #1 0 Gas #2 1001 Gas #3	N N	Calibrated
1/18/2013	TVV	Gas #1 0 Gas #2 1000 Gas #3	N N	Calibrated
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		

RESPONSE TIME CHECK[2]

Gas Standard SN: BAJ-289-990-1 Date: 1/17/2013 Conc: 1001

Time Response (sec.)	Instrument Response (ppm)	Comments
4	999	
4	1000	
4	1001	
Average Response Time (sec)	4.00	

[1] Calibration is required at the beginning of each sampling day.

[2] Response time check is required once per month.

FUGITIVE VOC EMISSION SURVEY
SCREENING REPORT

Second Quarter 2013

Norlite, LLC
628 South Saratoga Street
Cohoes, NY 12047

APPENDIX A

Microtip Calibration Data Sheets

40 CFR 60, APPENDIX A, METHOD 21
CALIBRATION PRECISION TEST DATA SHEET
(QUARTERLY)

Microtip Serial #: PPWM0051

#1 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc. (ppm):** 1001

#2 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

#3 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

DATE	INITIALS		METER READINGS	AVERAGE DIFFERENCE[1]	CALIBRATION PRECISION %[2]	COMMENTS
4/29/2013	TVV		#1 Gas-			
		a	995	5.33	0.53	
		b	995			
		c	997			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
4/30/2013	TVV		#1 Gas-			
		a	997	3.67	0.37	
		b	998			
		c	997			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
			#1 Gas-			
		a		1001.00	100.00	
		b				
		c				
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
			#1 Gas-			
		a		1001.00	100.00	
		b				
		c				
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			

[1] Average difference is done by calculating difference between each meter reading and certified concentration, divided by the number of meter readings.

[2] Calibration precision percentage is obtained by dividing value in step (1) by the certified concentration and multiplying by 100.

MICROTIP CALIBRATION DATA SHEET

Microtip Serial #: PPWM0051

#1 Gas Standard (type): Air, zero **SN:** 0129005-001 **Certified Conc.:** zero air

#2 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc.:** 1001 ppm

#3 Gas Standard (type): **SN:** **Certified Conc.:**

CALIBRATION DATE[1]:	CALIBRATOR	RESPONSE	RECALIBRATED?	COMMENTS
4/29/2013	TVV	Gas #1 0 Gas #2 999 Gas #3	N N	Calibrated
4/30/2013	TVV	Gas #1 0 Gas #2 1001 Gas #3	N N	Calibrated
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		

RESPONSE TIME CHECK[2]

Gas Standard SN: BAJ-289-990-1 Date: 4/29/2013 Conc: 1001

Time Response (sec.)	Instrument Response (ppm)	Comments
4	1000	
3	1000	
3	1001	
Average Response Time (sec)	3.33	

[1] Calibration is required at the beginning of each sampling day.
[2] Response time check is required once per month.

FUGITIVE VOC EMISSION SURVEY
SCREENING REPORT

Third QUARTER 2013

Norlite, LLC
628 South Saratoga Street
Cohoes, NY 12047

APPENDIX A

Microtip Calibration Data Sheets

MICROTIP CALIBRATION DATA SHEET

Microtip Serial #: PPWM0051

#1 Gas Standard (type): Air, zero **SN:** 0129005-001 **Certified Conc.:** zero air

#2 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc.:** 1001 ppm

#3 Gas Standard (type): **SN:** **Certified Conc.:**

CALIBRATION DATE[1]:	CALIBRATOR	RESPONSE	RECALIBRATED?	COMMENTS
7/15/2013	TVV	Gas #1 0 Gas #2 1001 Gas #3	N N	Calibrated
7/16/2013	TVV	Gas #1 0 Gas #2 1000 Gas #3	N N	Calibrated
7/17/2013	TVV	Gas #1 0 Gas #2 1001 Gas #3	N N	Calibrated
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		

RESPONSE TIME CHECK[2]

Gas Standard SN: BAJ-289-990-1 Date: 7/15/2013 Conc: 1001

Time Response (sec.)	Instrument Response (ppm)	Comments
3	998	
3	999	
4	1001	
Average Response Time (sec)	3.33	

[1] Calibration is required at the beginning of each sampling day.

[2] Response time check is required once per month.

40 CFR 60, APPENDIX A, METHOD 21
CALIBRATION PRECISION TEST DATA SHEET
(QUARTERLY)

Microtip Serial #: PPWM0051

#1 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc. (ppm):** 1001

#2 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

#3 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

DATE	INITIALS		METER READINGS	AVERAGE DIFFERENCE[1]	CALIBRATION PRECISION %[2]	COMMENTS
7/15/2013	TVV		#1 Gas-			
		a	995	4.00	0.40	
		b	997			
		c	999			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
7/16/2013	TVV		#1 Gas-			
		a	998	2.33	0.23	
		b	998			
		c	1000			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
7/17/2013	TVV		#1 Gas-			
		a	997	2.33	0.23	
		b	999			
		c	1002			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
			#1 Gas-			
		a		1001.00	100.00	
		b				
		c				
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			

[1] Average difference is done by calculating difference between each meter reading and certified concentration, divided by the number of meter readings.

[2] Calibration precision percentage is obtained by dividing value in step (1) by the certified concentration and multiplying by 100.

FUGITIVE VOC EMISSION SURVEY
SCREENING REPORT

Fourth Quarter 2013

Norlite, LLC
628 South Saratoga Street
Cohoes, NY 12047

APPENDIX A

Microtip Calibration Data Sheets

40 CFR 60, APPENDIX A, METHOD 21
CALIBRATION PRECISION TEST DATA SHEET
(QUARTERLY)

Microtip Serial #: PPWM0051

#1 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc. (ppm):** 1001

#2 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

#3 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

DATE	INITIALS		METER READINGS	AVERAGE DIFFERENCE[1]	CALIBRATION PRECISION %[2]	COMMENTS
10/8/2013	TVV		#1 Gas-			
		a	999	1.33	0.13	
		b	999			
		c	1001			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
10/9/2013	TVV		#1 Gas-			
		a	998	3.00	0.30	
		b	997			
		c	999			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
			#1 Gas-			
		a		1001.00	100.00	
		b				
		c				
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
			#1 Gas-			
		a		1001.00	100.00	
		b				
		c				
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			

[1] Average difference is done by calculating difference between each meter reading and certified concentration, divided by the number of meter readings.

[2] Calibration precision percentage is obtained by dividing value in step (1) by the certified concentration and multiplying by 100.

MICROTIP CALIBRATION DATA SHEET

Microtip Serial #: PPWM0051

#1 Gas Standard (type): Air, zero **SN:** 0129005-001 **Certified Conc.:** zero air

#2 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc.:** 1001 ppm

#3 Gas Standard (type): **SN:** **Certified Conc.:**

CALIBRATION DATE[1]:	CALIBRATOR	RESPONSE	RECALIBRATED?	COMMENTS
10/8/2013	TVV	Gas #1 0 Gas #2 1001 Gas #3	N N	Calibrated
10/9/2013	TVV	Gas #1 0 Gas #2 1000 Gas #3	N N	Calibrated
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		

RESPONSE TIME CHECK[2]

Gas Standard SN: BAJ-289-990-1 Date: 10/8/2013 Conc: 1001

Time Response (sec.)	Instrument Response (ppm)	Comments
4	999	
4	998	
4	1000	
Average Response Time (sec)	4.00	

[1] Calibration is required at the beginning of each sampling day.
[2] Response time check is required once per month.

FUGITIVE VOC EMISSION SURVEY
SCREENING REPORT

First QUARTER 2014

Norlite, LLC
628 South Saratoga Street
Cohoes, NY 12047

APPENDIX A

Microtip Calibration Data Sheets

40 CFR 60, APPENDIX A, METHOD 21
CALIBRATION PRECISION TEST DATA SHEET
(QUARTERLY)

Microtip Serial #: PPWM0051

#1 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc. (ppm):** 1001

#2 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

#3 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

DATE	INITIALS		METER READINGS	AVERAGE DIFFERENCE[1]	CALIBRATION PRECISION %[2]	COMMENTS
1/27/2014	TVV		#1 Gas-			
		a	999	2.00	0.20	
		b	997			
		c	1001			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
1/28/2014	TVV		#1 Gas-			
		a	998	2.00	0.20	
		b	1002			
		c	1003			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
			#1 Gas-			
		a		1001.00	100.00	
		b				
		c				
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
			#1 Gas-			
		a		1001.00	100.00	
		b				
		c				
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			

[1] Average difference is done by calculating difference between each meter reading and certified concentration, divided by the number of meter readings.

[2] Calibration precision percentage is obtained by dividing value in step (1) by the certified concentration and multiplying by 100.

MICROTIP CALIBRATION DATA SHEET

Microtip Serial #: PPWM0051

#1 Gas Standard (type): Air, zero **SN:** 0129005-001 **Certified Conc.:** zero air

#2 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc.:** 1001 ppm

#3 Gas Standard (type): **SN:** **Certified Conc.:**

CALIBRATION DATE[1]:	CALIBRATOR	RESPONSE	RECALIBRATED?	COMMENTS
1/27/2014	TVV	Gas #1 0 Gas #2 1001 Gas #3	N N	Calibrated
1/28/2014	TVV	Gas #1 0 Gas #2 1000 Gas #3	N N	Calibrated
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		

RESPONSE TIME CHECK[2]

Gas Standard SN: BAJ-289-990-1 Date: 1/27/2014 Conc: 1001

Time Response (sec.)	Instrument Response (ppm)	Comments
4	1000	
3	1000	
3	1001	
Average Response Time (sec)	3.33	

[1] Calibration is required at the beginning of each sampling day.

[2] Response time check is required once per month.

FUGITIVE VOC EMISSION SURVEY
SCREENING REPORT

Second Quarter 2014

Norlite, LLC
628 South Saratoga Street
Cohoes, NY 12047

APPENDIX A

Microtip Calibration Data Sheets

MICROTIP CALIBRATION DATA SHEET

Microtip Serial #: PPWM0051

#1 Gas Standard (type): Air, zero **SN:** 0129005-001 **Certified Conc.:** zero air

#2 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc.:** 1001 ppm

#3 Gas Standard (type): **SN:** **Certified Conc.:**

CALIBRATION DATE[1]:	CALIBRATOR	RESPONSE	RECALIBRATED?	COMMENTS
4/17/2014	TVV	Gas #1 0 Gas #2 1000 Gas #3	N N	Calibrated
4/18/2014	TVV	Gas #1 0 Gas #2 1001 Gas #3	N N	Calibrated
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		

RESPONSE TIME CHECK[2]

Gas Standard SN: BAJ-289-990-1 Date: 4/17/2014 Conc: 1001

Time Response (sec.)	Instrument Response (ppm)	Comments
4	1000	
3	1000	
3	1001	
Average Response Time (sec)	3.33	

[1] Calibration is required at the beginning of each sampling day.
[2] Response time check is required once per month.

40 CFR 60, APPENDIX A, METHOD 21
CALIBRATION PRECISION TEST DATA SHEET
(QUARTERLY)

Microtip Serial #: PPWM0051

#1 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc. (ppm):** 1001

#2 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

#3 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

DATE	INITIALS		METER READINGS	AVERAGE DIFFERENCE[1]	CALIBRATION PRECISION %[2]	COMMENTS
4/17/2014	TVV		#1 Gas-			
		a	999	2.67	0.27	
		b	998			
		c	998			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
4/18/2014	TVV		#1 Gas-			
		a	1000	1.00	0.10	
		b	999			
		c	1001			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
			#1 Gas-			
		a		1001.00	100.00	
		b				
		c				
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
			#1 Gas-			
		a		1001.00	100.00	
		b				
		c				
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			

[1] Average difference is done by calculating difference between each meter reading and certified concentration, divided by the number of meter readings.

[2] Calibration precision percentage is obtained by dividing value in step (1) by the certified concentration and multiplying by 100.

FUGITIVE VOC EMISSION SURVEY
SCREENING REPORT

Third QUARTER 2014

Norlite, LLC
628 South Saratoga Street
Cohoes, NY 12047

APPENDIX A

Microtip Calibration Data Sheets

MICROTIP CALIBRATION DATA SHEET

Microtip Serial #: PPWM0051

#1 Gas Standard (type): Air, zero **SN:** 0129005-001 **Certified Conc.:** zero air

#2 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc.:** 1001 ppm

#3 Gas Standard (type): **SN:** **Certified Conc.:**

CALIBRATION DATE[1]:	CALIBRATOR	RESPONSE	RECALIBRATED?	COMMENTS
7/21/2014	TVV	Gas #1 0 Gas #2 1001 Gas #3	N N	Calibrated
7/22/2014	TVV	Gas #1 0 Gas #2 1000 Gas #3	N N	Calibrated
7/23/2014	TVV	Gas #1 0 Gas #2 999 Gas #3	N N	Calibrated
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		

RESPONSE TIME CHECK[2]

Gas Standard SN: BAJ-289-990-1 Date: 7/21/2014 Conc: 1001

Time Response (sec.)	Instrument Response (ppm)	Comments
4	1001	
4	1000	
3	999	
Average Response Time (sec)	3.67	

[1] Calibration is required at the beginning of each sampling day.
[2] Response time check is required once per month.

40 CFR 60, APPENDIX A, METHOD 21
CALIBRATION PRECISION TEST DATA SHEET
(QUARTERLY)

Microtip Serial #: PPWM0051

#1 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc. (ppm):** 1001

#2 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

#3 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

DATE	INITIALS		METER READINGS	AVERAGE DIFFERENCE[1]	CALIBRATION PRECISION %[2]	COMMENTS
7/21/2014	TVV		#1 Gas-			
		a	1000	1.67	0.17	
		b	1002			
		c	998			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
7/22/2014	TVV		#1 Gas-			
		a	997	1.67	0.17	
		b	1001			
		c	1002			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
7/23/2014			#1 Gas-			
		a	999	2.33	0.23	
		b	999			
		c	998			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
			#1 Gas-			
		a		1001.00	100.00	
		b				
		c				
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			

[1] Average difference is done by calculating difference between each meter reading and certified concentration, divided by the number of meter readings.

[2] Calibration precision percentage is obtained by dividing value in step (1) by the certified concentration and multiplying by 100.

FUGITIVE VOC EMISSION SURVEY
SCREENING REPORT

Fourth Quarter 2014

Norlite, LLC
628 South Saratoga Street
Cohoes, NY 12047

APPENDIX A

Microtip Calibration Data Sheets

40 CFR 60, APPENDIX A, METHOD 21
CALIBRATION PRECISION TEST DATA SHEET
(QUARTERLY)

Microtip Serial #: PPWM0051

#1 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc. (ppm):** 1001

#2 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

#3 Gas Standard (type): _____ **SN:** _____ **Certified Conc. (ppm):** _____

DATE	INITIALS		METER READINGS	AVERAGE DIFFERENCE[1]	CALIBRATION PRECISION %[2]	COMMENTS
10/8/2014	TVV		#1 Gas-			
		a	1000	1.00	0.10	
		b	1001			
		c	1003			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
10/9/2014	TVV		#1 Gas-			
		a	999	1.33	0.13	
		b	1000			
		c	1000			
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
			#1 Gas-			
		a		1001.00	100.00	
		b				
		c				
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			
			#1 Gas-			
		a		1001.00	100.00	
		b				
		c				
			#2 Gas-			
		a	a			
		b	b			
		c	c			
			#3 Gas-			
		a	a			
		b	b			
		c	c			

[1] Average difference is done by calculating difference between each meter reading and certified concentration, divided by the number of meter readings.

[2] Calibration precision percentage is obtained by dividing value in step (1) by the certified concentration and multiplying by 100.

MICROTIP CALIBRATION DATA SHEET

Microtip Serial #: PPWM0051

#1 Gas Standard (type): Air, zero **SN:** 0129005-001 **Certified Conc.:** zero air

#2 Gas Standard (type): N-Hexane **SN:** BAJ-289-990-1 **Certified Conc.:** 1001 ppm

#3 Gas Standard (type): **SN:** **Certified Conc.:**

CALIBRATION DATE[1]:	CALIBRATOR	RESPONSE	RECALIBRATED?	COMMENTS
10/8/2014	TVV	Gas #1 0 Gas #2 1002 Gas #3	N N	Calibrated
10/9/2014	TVV	Gas #1 0 Gas #2 1001 Gas #3	N N	Calibrated
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		
		Gas #1 Gas #2 Gas #3		

RESPONSE TIME CHECK[2]

Gas Standard SN: BAJ-289-990-1 Date: 10/8/2014 Conc: 1001

Time Response (sec.)	Instrument Response (ppm)	Comments
3	1000	
3	1001	
3	999	
Average Response Time (sec)	3.00	

[1] Calibration is required at the beginning of each sampling day.
[2] Response time check is required once per month.